

SPECIFICATION

Electronic Version 1.2.8

Stylesheet Version 1.0

Ionic Pre-concentration XRF Detection and Analysis Device, System and Method

Background of Invention

[0001] This invention relates to the field of x-ray fluorescence analysis systems with pre-concentration devices, particularly for the in-situ measurement of ultra-trace levels of ionic contaminants in aqueous solutions.

[0002] Virtually all elements in the periodic table form compounds that are soluble in water. Dissolved impurities exist in aqueous streams as positive and negative charged pairs of anions and cations. Many of these impurities are toxic for human consumption or harmful to industrial processes. Therefore, the detection and quantitative measurement of the concentration of these impurities is of interest.

[0003] The present invention was developed to assist industries which have a continuous need to in-situ monitor effluent discharges to sanitary systems, and industries within the manufacturing arena where ultra pure fluids are required, such as for the production of microchips, or where fairly precise concentrations of trace elements within fluids are desired. This invention is also useful to ensure the quality and purity of intake or supply waters more generally to meet health and environmental standards. Current methodologies for the detection of trace materials, e.g., trace metals and other ionic components within process fluid streams, typically require samples to be prepared for off line analysis by specialized laboratories, which can be both costly and time consuming. Valuable production time is lost while waiting for results of required tests, or worse, when unacceptable concentrations are allowed to pass through the system for lack of continuous, on-line monitoring capability.

[0006] The sensitivity of XRF techniques depends upon the achievable signal to noise ratio. This is determined by the number of photons emitted from the element of interest at its characteristic x-ray energy that can be detected compared to photons of nearby energies scattered into the detector via other processes. Because the background is randomly distributed, the signal to noise ratio can be improved by extending the exposure time. A carefully constructed standard XRF unit is able to achieve sensitivity of about 1 to 10 parts-per-million (ppm) in mass ratio with about 30 minutes of exposure. However, to improve that sensitivity by a factor of two statistically requires a factor of four times as long a time, i.e., sensitivity improvement varies only with the square root of time. Therefore, to detect trace-level impurities at sub-ppm, the exposure times required would be inordinately long. Long exposure

time usually introduces other non-statistical limitations, such as, but not limited to, gain drift in the detector, source output stability, etc. Thus, for all practical purposes, real-time analysis by direct XRF technique is limited to the ppm level since any greater sensitivity requires exposure times in excess of a few hours.

[0007] For quantitative work at better (greater sensitivity) than the 1 – 10 ppm level, processing of the material by means such as leaching, filtration, surface treatments, etc., by a highly trained individual is required to control "matrix" effects. This can take from several hours to days. This lengthy time between collection and reporting of analytical results prohibits XRF from being considered a 'real time' analytical method at ppm or better sensitivity levels.

[0008] The detection sensitivity can be improved by pre-concentrating the sample. Various filtration techniques have been tried for the detection of impurities associated with particulate matters, and ion exchange and chelating membranes have been utilized for the pre-concentration of dissolved compounds. These techniques typically require complex reversal processes such as removal or cleansing of a filter or drastically changing the pH of an ion exchange bed, for example. The capability to perform these pre-concentration steps and the subsequent analyses *in-situ* generally requires intervention of highly trained personnel and is not easily automated for on-site, real-time measurements.

[0009] Technologies used for extracting trace materials such as trace metals and ionic components in flow streams are often based upon or incorporated into water purification technologies. For example, U.S. patents 5,954,937 and 5,425,858 both by Farmer disclose an electrochemical cell for the removal of dissolved impurity ions from a liquid medium for purification purposes. The invention makes use of a highly porous carbon aerogel with very high surface density to form the electrodes of a capacitor. Upon the application of a bias voltage, the dissolved ions are attracted to the respective electrodes where they are captured in a "double layer" structure. The process can be reversed electrically to regenerate the cell. The concept is similar to earlier patents by Andelman (U.S. Patent Nos. 5,192,432, 5,200,068, 5,360,540, 5,415,768) that make use of a carbon fiber as the porous material wherein the capacitor is wound in a spiral configuration for the fluid to flow through the

electrodes, whereas Farmer uses a stack of capacitors requiring the fluid to flow between the electrodes. Farmer also recognizes that upon regeneration, the ions captured in the double layer can be discharged to a secondary chamber with significantly higher concentration to facilitate detection (see US 5,954,937, column 27, line 43 through column 28, line 7). But, Farmer provides no disclosure on specific arrangements for in-situ or remote measurements, nor on how to configure and operate a cell for such measurements, nor on how quantitative information regarding the concentration of the impurities in the flow stream can be determined. Clearly Farmer has not considered all of the technical issues that need to be resolved in order to obtain quantitative, in-situ or remote measurements of trace level impurity concentration in a flow stream.

[0010] For detection of trace level impurities in flow streams, additional prior art is found in chemical based concentration systems. U.S. Patent No. 5,834,633 by Davison claims the use of a permeable membrane capable of binding the impurity as a sampling device to collect and concentrate metal ions to facilitate detection using laboratory equipment such as is used in proton-induced X-ray emission (PIXE) techniques.

[0011] The prior art described above would be improved upon by an apparatus and method which provides a pre-concentration device capable of providing fully automated in-situ or remote analysis in real-time to provide quantitative measurements of trace materials in a fluid matrix, across a broad range of elements and concentration levels, by using well established XRF techniques but with significantly greater sensitivity than is made possible using standard XRF techniques and measurements.

Summary of Invention

[0012] The invention disclosed herein comprises a pre-concentration cell that can be integrated with an x-ray fluorescence analysis system to achieve significantly improved sensitivity in the detection and measurement of trace element concentrations in flow streams (elements in fluids). The device, system and method herein disclosed is capable of being implemented on-line for fully automated, in-situ, quantitative measurements of these trace element concentrations in flow streams. The concentrations of these impurities are usually reported in parts-per-million (ppm) or

parts-per-billion (ppb) of mass units, or equivalently as microgram/gram for ppm, and nanogram/gram for ppb. At such low concentrations, the detection and quantitative measurement of these impurities is nontrivial, and requires novel and non-obvious devices, systems, and methods.

[0013] It is important at the outset to distinguish the *removal* of impurities from the *detection and measurement* of impurities. For the detection and quantitative measurement of impurity concentrations by X-ray fluorescence, where the objective is to obtain a quantitatively accurate concentration measurement, the design and operation of a concentration cell must be significantly different than for desalinization, or deionization more generally, where the objective is to saturate the electrode surfaces in order to maximally remove all dissolved ions.

[0014] In general, the ability to detect an element present in trace amounts is determined by the intensity of the characteristic signal above background noise. The signal intensity is directly proportional to the concentration of the element, whereas the background noise arises from all other sources. The pre-concentration cell extracts ultra-trace levels of ionic contaminants from aqueous solutions in order to bring the intensity of the signal above background noise, so it becomes detectable by XRF or similar techniques.

[0015] Because most dissolved impurities exist as charged pairs, they can be concentrated by flowing the liquid in which they are contained through an energized capacitor. This process is commonly known as "capacitive deionization." The charged particles are drawn toward the positive and negative electrodes at a rate determined by their respective mobility in the solution and are captured by the respective electrodes to form what is known as a double layer. The amount of charge that can be collected is determined by the surface area available on the electrodes. This fact is exploited in the capacitive deionization processes to remove dissolved impurities from water, such as those disclosed in the patents by Andelman (U.S. Patent Nos. 5,192,432, 5,200,068, 5,360,540, 5,415,768, 5,547,581, 5,620,597, 5,779,891), Farmer (U.S. Patent Nos. 5,425,858, 5,954,937) and Otowa (U.S. Patent No. 5,538,611).

[0016] If this flow were to be continued indefinitely, more and more of the anions and

cations would become attached to the surface of the electrodes. Eventually, saturation would occur when the total charge on the electrode equals the ratio of the capacitance of the capacitor over the applied potential. Once saturated, while there is no further net charge buildup, an ion exchange process would continue through which high valence state ions replace low valence state ions in the double layer. By allowing this process to continue, eventually, the high valence state ions would be preferentially captured by the double layers in the electrode. Ordinarily, the desired data is acquired long before saturation, though in some circumstances, saturation may be desirable.

[0017] Because the capacitance of a capacitor is proportional to the surface area available for collection of charges, electrodes made of porous, pervious material with large surface to volume ratio are capable of holding significantly more electrical charges before reaching saturation than non-pervious materials. For application as electrodes, the material must also be electrically conducting. One such material is carbon in the forms of fiber or porous foam. A type of carbon foam known as aerogel can achieve surface density as high as $10^3 \text{ m}^2/\text{gm}$, consisting of ultra-fine cells with average pore sizes of less than 50 nm. In general, the higher the density of pores and the smaller the pore size, the higher the surface density in area per mass unit. This aerogel material is extensively used as the electrode material for ultra-capacitors. Farmer's patent for desalinization based on the capacitive deionization process uses this aerogel material.

[0018] The basic invention comprises a single capacitor with two electrodes made of porous, low atomic number conducting material with very high surface density separated by a small gap to form a flow channel for the fluid being characterized. The pressure and flow rate of the sample fluid must be carefully regulated and optimized for purpose of enhanced *detection and quantification, rather than deionization*. With the application of a voltage to the electrodes, the dissolved ions are collected on the surface of the electrodes to build up the concentration needed for detection by XRF or similar technique. Upon shorting out the electrodes, all the captured ions are returned into the flow stream. The process, therefore, is reversible, and the preconcentration cells are reusable.

[0019] In general, the XRF technique requires a source to illuminate the sample with

primary X-rays. In the present case, the sample comprises the electrodes of the capacitor with the captured impurity ions. Upon excitation by the primary X-ray, the impurity ions will fluoresce with emissions of photons of their characteristic energy. A detector capable of resolving the energy of the photons must be properly situated to collect the emitted photons. The measured intensity in each energy channel is then related to the concentration of the impurity on the electrode. The procedure for deducing quantitative information on the concentration of the impurities in the flow stream from the measurements is an important part of this invention.

[0020] Because the electrode material is porous, it must be enclosed by a non-porous material of sufficient strength to contain the flow pressure. On the other hand, the X-ray source and the detector must be situated outside the flow enclosure. The presence of the enclosure attenuates both the incident primary x-ray and the emitted secondary fluoresce photons. This is not a problem when the purpose is removal, e.g., deionization only. But for *quantitative XRF analysis*, inattention to the enclosure design will significantly reduce the sensitivity of the XRF technique. Therefore, the enclosure must be designed with a window that is highly transparent to the fluoresce photons. The design of the window and selection of suitable material for the window is also an important part of this invention.

[0021] From signal to noise considerations, it is desirable that the sample being measured contain a minimum of intervening materials in the path of the primary X-ray and the secondary fluorescence photons. In the present case, the intervening materials includes the cell window for passage of the X-rays, the electrode material, and the contained fluid ("matrix") being tested. The design of the concentration cell to enhance the signal to noise ratio is also an important part of this invention.

[0022] As will be discussed, it is also important to extract only a small percentage of impurities from the flow stream, and the flow rate must be carefully established and controlled to achieve this objective. Thus, determining and controlling a suitable flow rate of the fluid through the preconcentration cell is also an important part of this invention.

[0023] The present invention is intended for utilization both at the source point for on-line installation and real-time measurements, as well as for measurements of samples

trace elements, that can achieve higher sensitivity across a broader range of low-concentration levels.

- [0027] In particular, it is desirable for such a system to have a support structure that allows better penetration by x-rays or other radiation used for analysis, that avoids the entraining of excess fluids which can hinder the full range and sensitivity of detection, and that uses electrode materials better suited to detection and measurement rather than to removal of trace material concentrations.
- [0028] It is also desirable to provide optimum flow rates and pressures for the purpose of detecting and measuring rather than removing trace elements.
- [0029] It is also desirable to provide a system which substantially replaces trace elements back into the flow stream from whence they came, once detection is complete.
- [0030] It is also desirable to provide an x-ray fluorescence analysis system with an integrated ionic pre-concentration device.
- [0031] It is also desirable to provide for a portable, in-situ, x-ray fluorescence analysis system with an integrated ionic pre-concentration device.
- [0032] It is also desirable to provide for a stand-alone concentration cell that can be used to sample a flow stream for remote, off-line utilization, wherein the cell can later be transported to an x-ray source for measurement and analysis of the impurity concentrations in the flow stream, which is again optimized for XRF measurement rather than removal.
- [0033] It is also desirable to provide for a method of analyzing multiple species dissolved ionic concentrations, for purposes of both species identification *and quantitative measurement of each species* , as desired.
- [0034] It is also desirable to provide for the ability to perform such an analysis by remote telecommunications utilizing technology such as wireless transmission or the Internet .
- [0035] It is also desirable to provide for a system which can be operated by lay personnel without advanced scientific training.

Brief Description of Drawings

- [0036] The features of the invention believed to be novel are set forth in the associated claims. The invention, however, together with further objects and advantages thereof, may best be understood by reference to the following description taken in conjunction with the accompanying drawings in which:
- [0037] Fig. 1 is a top view of a preferred embodiment of the pre-concentration cell.
- [0038] Fig. 2 is a front cross-sectional view of the preferred embodiment of the pre-concentration cell.
- [0039] Fig. 3 is an exploded two-dimensional side view of the electrode assembly of the pre-concentration cell.
- [0040] Fig. 4 is a front cross-sectional view of the assembled electrode assembly.
- [0041] Fig. 5 is a top view of the upper or lower high surface area electrode.
- [0042] Fig. 6 is a top view of the interelectrode spacer.
- [0043] Fig. 7 is a top view of the upper or lower sealing o-ring.
- [0044] Fig. 8 is a top view of the upper or lower contact ring.
- [0045] Fig. 9 is a top view of the upper or lower transmission window.
- [0046] Fig. 10 is a top view of the upper or lower window compression ring.
- [0047] Fig. 11 is a top view of the upper or lower electrical contact ring.
- [0048] Fig. 12 is a top view of the upper or lower retaining ring.
- [0049] Fig. 13 is a top view of the preferred embodiment of the pre-concentration cell including control modules.
- [0050] Fig. 14 is a schematic view of the overall XRF detection and analysis device, system and method.
- [0051] Fig. 15 is a front cross-sectional view of the operation of the pre-concentration cell.

- [0052] Fig. 16 is a schematic view illustrating the double layer formation in the ionic pre-concentration cell.
- [0053] Fig. 17 is a flow chart detailing the measurement and analysis system operation when conducted *during preconcentration*.
- [0054] Fig. 18 is a front cross-sectional view illustrating the cleaning of the pre-concentration cell.
- [0055] Fig. 19 is a top view of the pre-concentration cell with voltage application means comprising a transportable voltage supply for preconcentration without immediate analysis, enabling subsequent cell transportability to an analysis site.
- [0056] Fig. 20, in contrast to Fig. 17, is a flow chart detailing the measurement and analysis system operation when conducted *after preconcentration*.
- [0057] Fig. 21 is a top view of an alternative embodiment of the preconcentration cell.
- [0058] Fig. 22 is a top cross-sectional view of an alternative embodiment of the preconcentration cell with inlet flow tubes.
- [0059] Fig. 23 is a top view of an alternative embodiment of the pre-concentration cell with flow turbulence enhancers.

Detailed Description

[0060]

The invention disclosed herein comprises a pre-concentration cell that can be integrated with an x-ray fluorescence system and which is capable of being implemented as a fully automated system for in-situ installation and measurements. The pre-concentration cell extracts ultra-trace levels of ionic impurities from aqueous solutions in order to enhance their detection by x-ray fluorescence, and allows for the use of XRF techniques in the *sub-part-per-million* (sub-ppm) minimum detection level (MDL) region. The pre-concentration cell can be installed on-line for in-situ data collection, analysis, and automatic reporting without human intervention, and because the pre-concentration times are comparatively short, it can be employed in a real time system. The ability to detect such contaminants in real time, and to measure them accurately, is an important improvement over existing techniques. It can also be used

for remote sampling, and then transported elsewhere for irradiation and analysis.

[0061] This invention is applicable to the detection of a wide variety of ionic contaminants in a wide variety of fluid matrices. For example, this present invention is well-suited for detecting and measuring specific metal contaminants, such as arsenic, in drinking water, and for detecting and measuring other toxic metals such as lead and mercury in industrial discharge streams. It is also useful for detecting and measuring elements in other fluids, such as to ensure that fluids used for various industrial processes are as pure as is needed for those processes, or contain desired elements in desired concentrations.

[0062] A primary component of the invention disclosed herein is the ionic pre-concentration cell 100 of Figs. 1 and 13, (and the very similar transportable pre-concentration cell 101 of Fig. 19). Figs. 21 and 22 show some alternative cell embodiments. It is well understood that other embodiments are also feasible within the scope of this disclosure and its associated claims. This component will be described in detail in order to better understand the functionality of the in-situ XRF analysis system.

[0063] Pre-concentration cell 100 in its various embodiments has the ability to extract ultra-trace levels of ionic contaminants from aqueous solutions. In order to determine the concentration of an impurity in the flow stream, the ratio of the rate of concentration to the flow rate must be maintained constant. Because the rate of concentration is proportional to the concentration of the impurity in the flow stream, maintaining adequate flow rates through the cell 100 ensures that the basic concentration levels in the fluid matrix remain substantially unchanged, resulting in a highly accurate quantitative measurement. This ionic pre-concentration cell 100 provides means of detecting sub ppm levels of dissolved ionic contaminants, a sensitivity not heretofore achievable with XRF and similar analytical techniques outside of a laboratory-based system.

[0064] As shown in top view Fig. 1 and front view Fig. 2, the ionic pre-concentration cell 100 comprises a cell collector body 102 with an inlet flow port 104, an outlet flow port 106, an inlet flow slot 108, an outlet flow slot 110, an electrode assembly 112 (detailed further in Figs. 3 and 4), and tightening fasteners 109. Preferably, the cell

collector body 102 comprises a non-conductive, low-Z material suitable for use in a high radiation environment. In general, any material is suitable for cell collector body 102 that is non-metallic and non-conductive with the capability or characteristic of being readily machined into shapes. To be avoided are metals such as iron or copper that could leach dissolved ions into the flow stream and thereby cause a false reading (thus the material used must have suitable resistance to ionic leaching). The material must also be resistant to radiation degradation at the x-ray energies to which it will be subjected during testing. Delrin[®] plastic or any similar material with the aforementioned properties that is known in the art or may become known in the future is suitable for this purpose. Other possibilities include, but are not limited to, other plastics, fiberglass products such as NEMA G-10, or even glass so long as the glass utilized can be properly machined.

[0065] The inlet flow port 104 and outlet flow port 106 are preferably passageways laid out substantially parallel to each other and substantially on the same plane on opposite ends of the pre-concentration cell 100. Their illustrated circular cross-sectional shape in Fig. 2 can be varied at will within the scope of this disclosure and its associated claims so long as they permit the aqueous solution being tested to flow therethrough. Between the inlet flow port 104 and the outlet flow port 106, is electrode assembly 112, that connects the flow channel from inlet to outlet of the pre-concentration cell 100, as can be seen by examining Figs. 1 and 2.

[0066] Projecting substantially perpendicularly from inlet flow port 104 to electrode assembly 112 is an inlet flow slot 108 which lies in substantially the same plane as a central flow interelectrode gap 222 (see Figs. 3 and 4) of electrode assembly 112. Similarly, on the opposite side of the electrode assembly 112, an outlet flow slot 110 connects the central flow interelectrode gap 222 to outlet flow port 106. Outlet flow slot 110 is substantially perpendicular to the outlet flow port 106. These perpendicular configurations of the flow ports 104 and 106 relative to the flow slots 108 and 110 are non-limiting and for illustration only, and other configurations are equally acceptable within the scope of this disclosure and its associated claims.

[0067] Figs. 3 and 4 detail electrode assembly 112. High surface area electrodes 218 and 220 separated by central flow interelectrode gap 222, fixed by interelectrode spacer

231 or similar spacing means, comprise an electrical capacitor. Moving outward from central flow interelectrode gap 222, the upper and lower high surface area electrodes (218 and 220) are held in place by upper and lower contact rings 226 and 224, which is sealed by upper and lower sealing o-rings 223 and 225, as shown in a top view in Fig. 7 and 8, respectively. The contact rings 224 and 226 have essentially a stepped cross-section to hold the electrodes 218 and 220 and windows 228 and 232 in place. The cross-section, as specifically shown in Fig. 3, is a functional design but not in any way a limitation. To minimize noise, contact rings 224 and 226 are preferably fabricated from a low-Z carbon compound or compound with similar electrical and corrosion resistant characteristics.

[0068] An upper transmission window 228 highly transparent to the x-ray excitation and fluorescence radiation is held in place, essentially flush with the upper electrode 218 within the contact ring, by an upper window compression ring 219, an upper electrical contact ring 229 and an upper retaining ring 230. Similarly, an optional lower transmission window 232 is held in place by a lower window compression ring 221, a lower electrical contact ring 227, and a lower retaining ring 234. These transmission windows 228 and 232 seal electrode assembly 112 within the pre-concentration cell 100. Voltage is applied to electrodes 218 and 220 via voltage application means for applying a voltage differential. For example, without limitation, voltage is applied to the upper electrode 218 via the upper electrical contact 336 located on the upper electrical contact ring 229. Similarly, voltage is applied to the lower electrode 220 via the lower electrical contact 338 on the lower electrical contact ring 227. Alternatively, but again only as an example without limitation, retaining rings 230 and 234 may themselves double as the electrical contact to the electrodes 218 and 220. Indeed, any suitable voltage application means such as described above or otherwise known or which may become known in the art is acceptable within the scope of this disclosure and its associated claims.

[0069] When the liquid under analysis flows through the inter-electrode gap 222 with a voltage applied to the electrodes and thereby causing the electrodes to act as a capacitor, oppositely charged ions in the flow stream are attracted to the charged plates where they are trapped in a double layer formation. The amount of charge that the capacitor can hold is proportional to the total surface area of said capacitor. An

ideal material for such a capacitor to be used as a pre-concentration device for XRF analysis needs to have good conductivity, is porous with optimum pore sizes (large enough to facilitate ion transport in and out of the material and small enough to have a very high internal surface to void ratio), and is made of the lowest atomic number material possible (to be as transparent as possible to x-rays).

[0070] Nanocellular carbon (NCC), manufactured by Ocellus Inc. of Livermore, California, which is disclosed in U.S. Patent 5,945,084, is an ideal material for this purpose, though it is understood that other materials with similar properties that might be developed in the future may also be suitable as well within the scope of this disclosure and its associated claims. NCC comprises conductive sheets of carbon with very high surface to volume ratio by incorporating a distribution of pore sizes narrowly grouped around a well-controlled optimum size. The optimum pore size is selected to enable ions to readily migrate into the pores for double layer formation and out again for cleaning upon removal of the voltage.

[0071] While the aerogel employed by Farmer could also be used for pre-concentration in accordance with this invention, it is less preferred, since it has significantly more smaller pores that are too small to establish formation of double layers and inhibit the transport of the ions to and from the electrode surfaces. Trapped ions that are not removed during the self-cleaning cycle adds to a cumulative background and decrease the sensitivity of the XRF analysis over time.

[0072] Thus, for the concentration of dissolved ions to facilitate detection that is of interest here, the high surface area electrodes 218 and 220 preferably comprise NCC or any similar material that may be developed in the future in which the size of the pores can be optimized for double layer formation and for transport of the ions, and less preferably, but still within the scope of the invention, a carbon aerogel. For formation of double layers and ease of ion transport in and out of the pores, the characteristic dimension of the pores should be in the 20 ± 10 nm range. NCC with properties similar to what is disclosed within U.S. Patent 5,945,084 is preferred because it can be manufactured with a higher proportion of its pore sizes engineered to achieve high surface area with average pore size clustered around the desired value to provide greater collection capability for the desired ions, and for ease of cleansing.

[0073] The specific properties of this NCC electrode material which make it suitable for this application include a large plurality of pores characterized by a specific surface area of at least approximately $100 \text{ m}^2/\text{g}$; an average pore diameter of said pores between approximately 30 nm and 10 nm per pore; a distribution of pore diameters grouped with a standard deviation of less than approximately 10 nm around the average pore diameter; an x-ray transparency greater than approximately 90% for characteristic photon energies from an element of interest for which a fluidic concentration is to be measured by said system; electrical conductivity of 10–40 mOhms-cm when fabricated into a $\frac{1}{4}$ mm thick electrode; the ability to contain approximately at least 0.1% by weight of foreign material relative to the high surface area material prior to saturation; high structural rigidity wherein a displacement under the flow of the fluid does not exceed approximately 0.25mm; high wetting ability wherein an approximately $\frac{1}{4}$ mm thick sheet of the high surface area material becomes substantially wetted in less than approximately three seconds; and freedom from metallic impurities in excess of approximately .5 parts per million, when measured by XRF analysis. In order to achieve satisfactory performance, the use of any alternative or substitute material for the upper and lower high surface area electrodes 218 and 220 possessing similar properties to those outlined above is regarded to be within the scope of this disclosure and its associated claims.

[0074] Fig. 5 is a top view of the upper and lower high surface area electrodes 218 and 220. As shown in the figure, they are typically round in shape but could be formed in any alternative shape within the scope of this disclosure and its associated claims. Upper and lower high surface area electrodes 218 and 220, again, preferably comprise the NCC materials such as disclosed in U.S. Patent 5,945,084. The thicknesses of the high surface area electrodes 218 and 220 are determined by the transmission probability of the x-ray radiation to be used for analysis. For example, assuming that water is the liquid to be analyzed, then: if the radiation being analyzed is 15keV (typical of K-line radiation in the region of atomic number 39 or L-line radiation in the region of atomic number 95), the electrode material should be about 8 ± 2 mm thick; 10keV radiation (typical of K-line radiation in the region of atomic number 32 or L-line radiation in the region of atomic number 80) would require 2.5 ± 1 mm optimum thickness; while x-rays at or below 5keV (typical of K-line radiation in

the region of atomic number 23 or below or L-line radiation in the region of atomic number 59 or below) could only effectively transit less than ½ mm of electrode plus matrix (again assuming water for the matrix to be analyzed).

[0075] More generally, when an x-ray passes through matter, its intensity is attenuated exponentially as a function of the distance characterized by the mass absorption coefficient of the matter, μ (in cm^2/gm), which is a function of the photon energy and the atomic number of the material, see, for example, the Center for X-ray Optics X-Ray Data Booklet PUB-490, Douglas Vaughan ed. Lawrence Berkeley Laboratory, Oct. 1985, pages 2-28 through 2-48. In general, the lower the energy of the photon, and the higher the atomic number of the material, the larger is the value of the mass absorption coefficient. The numerical inverse of the product of the mass absorption coefficient (μ , in cm^2/gm) and the density of the material (ρ , in grams/cm^3) is given by $(\text{cm}) = 1/\mu\rho$, and is the optical depth of the photon in the material, that is, the path length of material of density ρ with mass absorption coefficient μ that will absorb $1/e$ or about 63% of the incident photons. Because of the attenuation effect, most of the fluorescence photon signals detected will have originated from the layer of the electrode on the order of the optical depth closest to the detector. Any additional materials present will contribute more significantly to the background than the signal. For this reason, to obtain maximum sensitivity, the thickness of the electrode nearer to the x-ray source and fluorescence detector should not exceed the optical depth of the photon to be detected, that is, this thickness should less than or equal to The two electrodes need to be matched (to $\pm 10\%$ or so) in order to maintain the desired electrical properties.

[0076] Thus, for example, for detection of arsenic, which has an x-ray energy at about 10 keV, the NCC electrode thickness can be as much as 1 mm, and even up to 2 mm. However, for detection of the lighter elements ($Z < 20$) for which the fluoresce photon energy is only a few keV, the electrode thickness should optimally be around 100 micron.

[0077] From the same signal to background noise consideration, the central flow interelectrode gap spacing 222 (predetermined interelectrode gap width) between electrodes 218 and 220 should be made very small to limit the amount of fluid

entrained between them, without excessively impeding the flow of the sample fluid through the cell. Excessive flow impedance would require a high pressure to induce the flow which would cause the thin window that is part of the system (described below) to bow out, and that would have detrimental effects. In special cases the Venturi effect could be utilized to balance the pressure required against the desired flow rate, however, this would limit the particular design to a limited range of flow rates depending upon the viscosity of the liquids being analyzed. Typically, a characteristic interelectrode gap spacing of approximately 2 mm is preferred. Minimum values range as low as 1 mm, .5 mm, and even .25 mm. Maximum values range as high as 2 mm, 5 mm and even 10 mm. The quantitative determination of an optimum width for the interelectrode gap is discussed in further detail later on.

[0078] A top view of the upper and (optional) lower transmission windows 228 and 232 is shown in Fig. 9. The transmission window is preferably round (but it, and all the other components in Figs. 5-11, can be any suitable shape within the scope of the disclosure and claims) and is fitted to integrate with the contact rings 224 and 226 and the window compression rings 219 and 221. Fig. 10 is a top view showing the details of the upper and lower window compression rings 219 and 221. As shown in these figures, the L-shaped ringed structure has a central opening to allow access to the transmission windows 228 and 232 while firmly holding them in place around their perimeter. Figure 11 shows a top view of the electrical contact rings 227 and 229 and Figure 12 shows a top view of the retaining rings 230 and 234. These rings are sized to include pass throughs to allow access to the transmission windows 228 and 232 and comprise indentations to allow for use of the tightening screws 109.

[0079] Upper and lower thin transmission windows 228 and 232 serve dual purposes by enabling the excitation x-ray and fluoresce photons to pass therethrough with minimum attenuation and also by acting as a fluid barrier to seal pre-concentration cell 100. Preferably, transmission windows 228 and 232 comprise Kapton®, a patented proprietary material manufactured by DuPont Chemicals. Kapton® is a hydrocarbon-based rigid plastic material that is ideal for x-ray radiation transmission (it is typically more transparent and scatters less radiation than most fluids of interest for analysis, by a factor of about ten) and its mechanical properties minimize window bowing (bending) so that less than a tenth of the thickness of liquid being analyzed is

entrained between the window and the electrode. Such bowing would be undesirable in this application because it would lead to fluid entraining between the bent window and the electrode that would then contribute to the attenuation of the X-ray and add to the background noise signals.

[0080] Any suitable material can be utilized for the x-ray transmission window as long as it is compatible with the fluid being analyzed, transmits x rays of interest efficiently (high signal), minimizes scattering of x rays that are not of interest (low noise), and is rigid enough, even as a thin window, to resist bending and thus remain in intimate contact with the high surface area electrodes 218 and 220 under any flow rates and pressures that might be used. Kapton is preferred for its mechanical stiffness compared to other suitable materials such as polypropylene, Saran, Formvar, Mylar, or Kimfoil, but it is understood that the use of these latter and similar materials also falls within the scope of this disclosure and its associated claims. Kapton is also preferred because it is low in contaminants as compared to Beryllium and compatible with a wide variety of fluids as compared to Boron Nitride. But, it is to be understood that other materials such as Beryllium and Boron Nitride could be used in place of Kapton if future advances in materials technology should manage to overcome these problems and if these materials can be fabricated as thin windows providing a water barrier with suitable x-ray transparency, corrosion resistance, and resistance to bowing.

[0081] As a more general rule, any material known or which may become known in the art is acceptable to be used for window 228 and optionally window 232, so long as that material comprises: an atomic number below 10; structural rigidity to support up to 1/10 atm. of pressure without bowing more than approximately 100 microns; substantial impermeability relative to said fluid; x-ray transparency greater than 90% for characteristic photon energies from an element of interest for which a fluidic concentration is to be measured by said system; x-ray scattering therefrom minimized to less than approximately 10% of radiation scattered from a column of said fluid equal to one optical depth in said fluid of a characteristic photonic energy from an element of interest for which a fluidic concentration is to be measured by said system; and freedom from any single contaminant in excess of 1 part per million, when measured by x-ray fluorescence.

[0082] Further, and also importantly, the Kapton or similar-material thin window is configured to be *in intimate contact with the high surface area electrode* . "Intimate contact," as used herein, is functionally defined to mean that the thin window is close enough to the high surface area electrode, and remains close enough to the high surface area electrode even under pressure from the flow of the fluid matrix under analysis, such that the absorption and scattering due to intervening entrained fluid between thin window and the high surface area electrode is less than approximately the absorption and scattering due to the window material itself.

[0083] The thickness of the window should be significantly less than the attenuation length for the energy of the x-ray yet thick enough to support the water pressure. For x-ray energies in the 1 to 10 KeV range for measurement of light elements, a preferred thickness for the thin Kapton windows is approximately 8 microns, to maintain maximum sensitivity for the measurement of light elements. For measurements of heavier elements, the x-ray energy needs to be higher in the 10–20keV range and a thicker window (8–25 microns) may be used to allow for higher pressure to increase the flow rate.

[0084] The cross-section of the pre-concentration cell *100* shown in Fig. 4 and the exploded view in Fig. 3, detail the layout of the electrode assembly *112* within the pre-concentration cell *100* . Although the pre-concentration cell *100* is presented as a horizontal assembly throughout this disclosure, it is to be understood that it is orientation independent, such that it can be implemented in any configuration, such as horizontal or vertical, or anything in between, that fits best with the application. It is further observed that the terms "upper" and "lower" are used to refer to such elements of preconcentration cell *100* as the electrodes *218* , *220* and the windows *228* , *232* . These terms are to be understood not in terms of "higher" or "lower" with respect to a gravitational field, but in terms of how preconcentration cell *100* is to be oriented relative to the x-ray source means *644* and fluorescence detector *648* for XRF analysis. The "upper" preconcentration cell elements are to be understood as those that ultimately are to face the x-ray source means *644* and x-ray fluorescence detector *648* during the XRF analysis, and that the "lower" elements are those on the far side of this x-ray equipment. If the x-ray equipment were to be placed gravitationally below the cell, then the "upper" elements would be gravitationally lower

than the "lower" elements.

[0085] It is also to be noted that the various window properties discussed above are optional for the "lower" window, insofar as it is possible to simply seal the "lower" face of the preconcentration cell with any material that provides a suitable fluid barrier and will not degrade under x-ray exposure or contact with the fluid of interest. The x-ray scattering and transmission properties of the lower face of the cell are of much less importance than those of the "upper" window since it is the upper window through which the x-rays are transmitted and XRF readings are taken.

[0086] Figure 13 shows a top view of the preferred embodiment of the pre-concentration cell 100 including the necessary control modules. To enable automated operation of the system, the preconcentration cell is fitted with an inlet automated valve 380 connected to the inlet flow port 104 and an outlet automated valve 382 connected to the outlet flow port 106. Such valves are capable of being remotely controlled to turn on and off the flow stream into the pre-concentration cell. In addition, and desirably, ionic pre-concentration cell 100 comprises a variety of diagnostics to set, monitor and control system performance, such as, but not limited to: flow control means 384 comprising, for example, not limitation, a pressure sensor to set the flow potential and flow meter to monitor the flow rate; voltage application means 389 comprising, for example, not limitation, an electric potential and current meter 386 to set and monitor the applied potential (voltage differential); and leakage current monitoring means 388 to monitor the leakage current which gives a measure of the conductivity of the sample fluid due to presence of dissolved ions and the rate that they are being extracted from the flow stream.

[0087] The pre-concentration cell 100, as described above, is an enabling component to facilitate automated detection and measurement of impurities in fluids at sub-ppm concentration level as shown in Fig. 14. For purpose of illustration, the size of the pre-concentration cell is greatly exaggerated. As shown in Fig. 14, an x-ray source means 644 for providing energy for fluorescing the captured ions in pre-concentration cell 100 is situated at an angle above pre-concentration cell 100. X-ray source means 644 may also comprise an x-ray emitting, sealed radioactive source. As is common practice in XRF systems, a filter ladder (not shown) may also be included

between the X-ray source and the pre-concentration cell. The filter ladder holds filters of different material and thickness to shape the spectrum of the excitation X-ray from the source.

[0088] A beam stop assembly means 646 for confining the radiation to the pre-concentration cell 100 and preventing radiation leakage to the outside is placed opposite the x-ray source in the pre-concentration cell 100 lower window area. The beam stop assembly preferably comprises three elements, namely, silicon, copper and tin, with the silicon layer closest to the source/detector area and the tin layer furthest away. All beam stop assemblies known or which may become known in the art are considered as possible for use here.

[0089] Detector means 648 for collecting the fluorescence produced by the captured ions in pre-concentration cell 100 (i.e., for detecting photons emitted from preconcentration cell after it is irradiated) is positioned on the same side of the pre-concentration cell 100 as the the x-ray source 644 and is oriented for such collection. This arrangement is for purpose of illustration as one possible implementation. Any other arrangement that directs the x-ray to the pre-concentration cell and allows the fluoresce photons to be collected by the detector are also within the scope of this invention as disclosed and claimed.

[0090] Analysis means 652 , 654 for analyzing this fluorescence data, such as but not limited to the illustrated multi-channel analyzer and host computer, is used for processing and analyzing the fluorescence image data captured by x-ray fluorescence detector 648 . The host computer illustrated by 654 can optionally be multi-functioned to control the operation of the system and for data analysis.

[0091] The present invention utilizes as x-ray source means 444 , preferably, but without limitation, a Kevex model 5039S X-ray source which generates up to 50 keV electrons at 1.0 mA. The resulting Bremstrahlung radiation from a tungsten target generates the primary x-ray beam from the source. The input x-ray flux must have sufficient energy and intensity to excite the metal ions captured in the pre-concentration cell 100 to fluoresce in order for detection to take place. Alternatively, the present invention could also utilize as x-ray source means 644 , a sealed x-ray source, obtainable from vendors such as Isotope Sciences of Canoga Park, California. The advantage of sealed

sources is that they do not require power inputs from external sources. Similarly, any other type of x-ray source means 644 known in the art or which may become known in the art, which meets the functional requirements specified herein, is also suitable for use within the scope of this disclosure and its associated claims.

[0092] The preferred, but without limitation, x-ray fluorescence detector means 648 is an Amptek model XR-100CR solid state x-ray detector with 5mm crystal in combination with a multi channel analyzer (MCA8000) also from Amptek. This system is either battery powered or powered from conventional power sources. X-ray fluorescence detector means 648 may also comprise any other suitable detectors known or which may become known in the art, within the scope of this disclosure and its associated claims. A combined x-ray source and a detector as a single unit can also be used in conjunction with the concentration cell installed in the flow stream as a complete system for detection of impurity concentration in flow streams.

[0093] For fully automated operation, the system needs to be connected to a computer, such as a personal computer (PC). This can be separate from host computer illustrated as 654, or as noted above, can be a function provided by the host computer. The function of the computer is to control the operation of the system; to collect, store and analyze the data; and if desired, to transmit the data via a modem, wireless, or similar telecommunications means.

[0094] The functions required to control the operation of the computer may include but are not limited to operations which: turn on and off the valves to allow the sample flow stream through the system; set the pressure on the flow pressure regulator; monitor the outputs from the flow meter and regulate the flow pressure to maintain constant flow; set and apply a voltage to the electrodes; monitor and store the leakage current across the capacitor; turn on and off the XRF source if a non-radioactive source is used and set the voltage applied; turn on and off the multi-channel analyzer and record the output spectrum of the analyzer at frequent intervals; and accept inputs from an operator to set and store the desired operating parameters.

[0095] The functions required to analyze the data may include but not be limited to: analysis of the stored data to determine the count rate in each energy channel; detection of signals above background; and comparison of detected signals with a

stored lookup table to identify the elements, using a variety of commercially available or custom-developed algorithms.

[0096] The functions required for transmission of data may include but are not limited to connecting to a modem or transmitter or other telecommunications means at user specified-times or intervals, or to send a signal or trigger and alarm when the concentration of any specified impurities identified exceeds preset threshold values; and calling up the data and transmitting the data. Alternatively, all or some of these functions noted in this paragraph may be performed by human intervention.

[0097] Referring to Fig. 15, the operation of the XRF analysis system will now be described. Because an x-ray of sufficient energy is capable of inducing fluorescence in all elements, the materials comprising the cell itself will contribute to the signals detected. Therefore, prior to its use, each pre-concentration cell must be characterized to establish the background that it contributes, as well as its other pertinent characteristics. Each cell thus comprises *background data*, *sensitivity data*, and *ion extraction rate data* acquired from and associated therewith, which is measured after the cell is fabricated and assembled following the manufacturing process. This establishes a calibration for the cell.

[0098] To characterize the *background data* associated with the pre-concentration cell, the flow channel volume 222 is first filled with distilled water of high purity (or any other fluid to be analyzed, in a highly pure form). The distilled water is introduced through the inlet valve 380 of inlet flow port 104 until the cell is full of water. Then, the outlet valve 382 of outlet flow port 106 is shut off in order to trap the water in the cell body 100. The X-ray source 644 and the multi-channel analyzer 652 and detector 648 are then turned on when the cell is completely filled to record the signals received. The signals are photons of different energies arriving at the detector 648. The multi-channel analyzer 652 resolves the arriving photons into discreet energy channels according to their energy, to output a spectrum of intensities proportional to the number of photons in each energy channel detected. As the photons continue to arrive, the intensity in each energy channel will grow as a function of time.

[0099] This intensity spectrum consists of both a continuum due to scattering of the primary x-ray to different energies, and discreet bands due to photons from x-ray

fluorescence with energies characteristic of each element of the cell. The quantity of interest is the rate of growth of the intensity of each energy channel, which represents the rate that the fluorescence photons are generated and is a measure of the concentration of that particular element present. By sampling the intensity spectrum at frequent intervals, it is possible to obtain a plot of the intensity in a particular energy channel of interest as a function of time. The data points can be fitted by an analytical function by various mathematical techniques such as by least-mean-square method. The first derivative of this function is the rate at which photons of the particular energy are detected, which gives a relative measure of the concentration of a particular element in the cell.

[0100] Because the elements present in the cell are fixed, the rate of photons emitted, on average statistically, is constant. Therefore, if all operating parameters are maintained constant, the intensity of photons detected in each energy channel is expected to be a linear function of time, and its first derivative would be a constant. Determination of the rate of detection of photons in specific energy channels of interest establishes the background characteristics of the preconcentration cell.

[0101] In sum, this background data comprises data related to a rate at which photons are detected to be emitted from at least one background data energy channel of the preconcentration cell when the cell is filled with a highly purified form of a fluid of interest and exposed to x-rays.

[0102] Once the background is established, the concentration of impurities in any flow stream can be determined by observing the rate of increase of photons detected as the impurity ions are captured in the double layer when a voltage is applied to the electrode. However, in order to obtain quantitative measurement of the concentration of each element present, two additional calibration steps are needed. The first is to determine the sensitivity of the system in terms of relating the photon detection rate to the number of impurity elements present (sensitivity data). The second is to relate the rate at which the impurity elements are captured in the double layer to the concentration in the flow stream (ion extraction rate data). Both calibration steps are also part of the data associated with each concentration cell, and are also established following the manufacturing process.

[0103] The *sensitivity data* is obtained by replacing the distilled water of high purity with a first calibration solution containing known concentrations of one or more impurity ions of interest at levels *above the minimum detection levels* of the x-ray system. The procedure for establishing the background is then repeated. Because the concentrations of the specific impurities are above the MDL, the intensity of photons corresponding to their characteristic energy will grow at a faster rate than in the previous case. However, because their concentration is constant, the rate of growth will also be constant. Using the same procedure as before, it is then possible to obtain a rate of growth of the intensity. After subtracting out the growth rate due to the background, the result is the growth rate due to the presence of the impurities in the calibration fluid. Since the concentration of the impurity in the calibration fluid is known, it is now possible to quantitatively relate the growth rate to the impurity concentration quantitatively.

[0104] In sum, this sensitivity data comprises data related to a rate at which photons are detected to be emitted from at least one sensitivity data energy channel of the preconcentration cell when the preconcentration cell is filled with a first calibration solution containing at least one element of interest in the fluid of interest in known concentration above a *minimum detection level* of the system and exposed to x-ray.

[0105] The final calibration step is to relate the growth rate to the concentration in the flow stream, to obtain *ion extraction rate data*. For this step, the first calibration solution containing known impurities with concentration above MDL is purged from the system and a second calibration solution is spiked with one or more impurities of interest at known concentration levels *below the minimum detection levels* of the system. The impurities are introduced into pre-concentration cell 100 under regulated pressure to maintain constant flow rate and are shown as input fluid flow 840 in Figure 15. Input fluid flow 840 is introduced to the pre-concentration cell 100 via the inlet flow port 104 and the inlet flow slot or tube array (for the tube array, see Fig. 20 discussed below). The input fluid flow 840 continues through the central flow interelectrode gap 222 of the electrode assembly 112, and is exposed to a predetermined voltage applied across the upper and lower high surface area electrodes 218 and 220, respectively, via the upper and lower electrical contacts 336 and 338, respectively. Importantly, this predetermined voltage is set below the

electrochemical potential of the impurity ion species in the calibration fluid as well as the potential that can lead to electrolysis of water so as to avoid any permanent changes in the background readings as a result of performing measurements.

[0106] The presence of the applied voltage will cause the ions in the flow stream to migrate towards the oppositely charged electrode. The rate of migration of each ionic species is determined by its electric mobility in the fluid. As the ions reach the electrode, they are captured in the double layer, gradually building up their concentrations on the surface of the electrodes. During this time, the spectrum of photons cumulatively detected are continuously recorded at frequent intervals. The double layer formed between the upper electrode 218 and lower electrode 220 is detailed in Figure 16. The figure shows an exploded view of the interelectrode gap 222, the flow stream 840 through it, and the ion migration during the double layer formation.

[0107] The intensities of photons in the energy channels corresponding to the characteristic energy of the impurities in the calibration fluid is sampled at frequent intervals to determine their rates of growth. As long as the concentration of the impurity in the double layer is below the detection level of the instrument, the rate of growth will remain the same as for the background case and its intensity will remain as a linear function of time. With continued accumulation, however, the concentration will eventually exceed the minimum detection limit for the impurity. As a result, the intensity as a function of time will gradually deviate from being linear and the first derivative of this function will no longer be a constant, since the rate of detection is increasing. The second derivative of this function now provides the rate that the impurity ions are being captured in the double layer. Using the scaling factor that relates the count rate to the impurity concentration established for the cell, the rate can be related to the number of impurity ions being captured by the double layer. Since the concentration of the impurity in the flow stream is known, a second scaling factor that relates the growth rate to the concentration is established.

[0108] In sum, this ion extraction rate data (scaling factor) comprises data related to a rate at which photons are detected to be emitted from the at least one ion extraction rate data energy channel of the preconcentration cell when a second calibration

procedure for relating the growth of the count rate to the impurity in the calibration fluid spiked with a low concentration of the impurity of interest. Continuing with the description of Fig. 15, an electrostatic potential (voltage differential) applied to the upper and lower high surface area electrodes 218 and 220 and across the fluid flow 840 through central flow interelectrode gap 222 establishes a current of migrating ions from the fluid flow. For a specific gap 222 and a fixed applied voltage between the upper and lower high surface area electrodes 218 and 220, the current will self-adjust to match the conductivity of the sample to be analyzed.

[0112] Since the concentration of an impurity in the flow stream is determined by measuring the *rate* at which the impurity ions are captured in the double layer, it is essential that the concentration in the flow stream remain relatively constant. Therefore, *unlike the design for a deionization cell where the objective is to maximally extract the dissolved ions from the flow stream, the operation of this pre-concentration cell requires that only a small percentage of each impurity species present be extracted*, preferably not more than a 1% extraction percentage, with 2%, 3%, 4% and at most 5% being possible, but successively-less desirable options. This is accomplished by adjusting the flow rate to assure there is an adequate supply of the impurity ions in the flow stream to allow only a small percentage to be extracted in passing through the pre-concentration cell.

[0113] The migration of the dissolved ions in the flow stream to the electrodes appears as a leakage current j in the capacitor which is the sum of the partial currents j_i associated with each impurity ion of species i , given by

$$j = \sum_i j_i = \sum_i \mu_i n_i \frac{\Phi}{d} = \sigma \frac{\Phi}{d}$$

[0114] where μ_i is the mobility of the impurity ion in the matrix and n_i is its number density, and σ is the composite conductivity of the fluid, Φ is the potential applied across the electrodes separated by the distance d . The partial current associated with each ion species is proportional to its number density in the matrix. At very low concentration, it is important to ensure that the impurities are not significantly depleted in order to obtain accurate measurements of the impurity content.

[0115] The rate that the impurity ions are extracted per unit time is given by:

$$\Gamma_{ie} = \frac{j_i}{q} A$$

- [0116] where $q = 1.60 \times 10^{-19}$ Coulomb is the unit charge and A is the ordinary (as opposed to the high specific) surface area covered by the electrode. For purpose of illustration, consider a flow stream that contains only a single species of impurity, therefore,

$$\Gamma_{ie} = \frac{j}{q} A = \frac{\sigma \Phi}{qd} A = \frac{\mu_i n_i \Phi}{qd} A$$

- [0117] i.e., the single impurity is the sole contributor to the conductivity of the matrix.

- [0118] The rate at which the impurity ions are available in the flow stream for extraction is determined by the flow rate F in volume per unit time, and the concentration of the particular impurity. Concentration is defined by

$$C = \frac{n_i w_i}{n_f w_f}$$

- [0119] where n_i and n_f are the number densities of the impurity ion and of the matrix, and w_i and w_f are the atomic or molecular weights of the impurity and the carrier fluid, respectively.

- [0120] Hence, the rate at which the impurity ions are supplied to the cell is given by:

$$\Gamma_{is} = F n_i = F \frac{w_f}{w_i} C n_f$$

- [0121] Therefore, the percent of impurity extracted from the flow stream is given by:

$$\varepsilon = \frac{\Gamma_{ie}}{\Gamma_{is}} = \frac{\sigma \Phi w_i}{qd w_f n_f C F} A \times 100\%$$

- [0122] which suggests at first glance that the percent extraction is inversely proportional to the concentration.

- [0123]

However, very importantly, for water with a low concentration of dissolved ions in the ppb range, the conductivity is essentially a linear function of the concentration of the dissolved ions, with typical values in the range of 2×10^{-9} /Ohm-cm-ppb, i.e. the conductivity may be approximated as:

$$\sigma \approx 2 \times 10^{-9} \text{ C / Ohm-cm}$$

[0124] where C is in units of ppb. Substituting for the conductivity σ into the expression for ϵ , we find that the C terms cancel out, and thus that *at low concentrations the percent of impurity extracted from the flow stream is independent of the concentration.*

[0125] Thus, at sub-ppm concentrations:

$$\epsilon = \frac{\Gamma_{ie}}{\Gamma_{is}} = \frac{\sigma \Phi w_i}{q d w_f n_f C F} \times 100\% \approx 2 \times 10^{-9} \frac{\Phi w_i A}{q d w_f n_f F} \propto \frac{\Phi A}{d F}$$

[0126] It is also worth noting that an optimal size for the interelectrode gap, which was discussed earlier, is specified in terms of the above as:

$$d = \frac{\sigma \Phi w_i}{q \epsilon w_f n_f C F} \times 100\% \approx 2 \times 10^{-9} \frac{\Phi w_i A}{q \epsilon w_f n_f F} \times 100\% \propto \frac{\Phi A}{\epsilon F}$$

[0127] This is an important result for the quantitative determination of the concentration of elements in the flow stream, because it means that the calibration that relates the rate of build up of impurities on the electrodes to the concentration in the flow stream remains constant for all ranges of concentration of interest.

[0128] Differently stated, returning now to the detection of *more than one species of impurity*, this very important approximation means that so long as all the concentrations are in the range where conductivity varies substantially linearly with concentration, with variability governed predominantly by the atomic weight of the impurity species in question, a single flow rate selected to achieve an approximate 1% extraction rate for one species of impurity will also achieve an approximate 1% extraction rate for other species of impurity in this linear variation range, *even if the concentrations of the different species are substantially different from one another within that linear variation range.*

[0129] As a specific example illustrating the foregoing equations, consider Arsenic with atomic weight of 75 amu in a 1 ppb concentration flowing through the preconcentration cell with two electrodes each covering an ordinary surface area of 1 cm². To avoid any electro-chemical process from occurring, the applied potential needs to be kept below 1 volt. To minimize the amount of water entrained in the cell,

the gap spacing should be kept small. From practical considerations, a typical gap spacing might be 1 mm, though the discussion just preceding provides further detail regarding the quantitative optimization of this gap spacing.

[0130] The Arsenic ions will then be extracted at the rate on the order of:

$$\Gamma_{ie} = \frac{(2 \times 10^{-9} / \Omega - \text{cm})(1 \text{ Volt})}{(1.6 \times 10^{-19} \text{ Coulomb})(0.1 \text{ cm})} \approx 10^{11} \# / \text{sec}$$

[0131] To keep the extraction at 1% level, the cell needs to be supplied with 10^{13} ions/sec. Hence, the number of water molecules containing the Arsenic impurity at 1 ppb concentration that need to be supplied to the cell to maintain extraction at 1% level is given by:

$$n_f = \frac{w_i n_i}{w_f C} = \frac{75 (10^{13} \# / \text{sec})}{18 \cdot 10^{-9}} = 4.17 \times 10^{22} \# / \text{sec}$$

[0132] where the molecular weight of the matrix is taken as that of water with a value of 18 amu. The molecular density of water at room temperature is $3.3 \times 10^{22} \# / \text{cm}^3$. Hence, the flow rate needed is on the order of:

$$F = \frac{(4.17 \times 10^{22} \# / \text{sec})}{(3.35 \times 10^{22} \# / \text{cm}^3)} = 1.25 \text{ cm}^3 / \text{sec} = 1.25 \text{ mL} / \text{sec} = 74.7 \text{ mL} / \text{min}$$

[0133] Assuming the flow channel has a cross-sectional area of 0.1 cm^2 , the flow velocity required would be in the range of 12.5 cm/sec.

[0134] In contrast, suppose the arsenic is now in a 10 ppb (i.e., a 10^{-8}) concentration ten times as much as in the prior example. Arsenic ions will then be extracted at 10^{12} #/sec, requiring a 10^{14} #/sec supply. With w_i / w_f remaining constant, n_i / C goes from $10^{13} / 10^{-9}$ to $10^{14} / 10^{-8}$, and n_f remains constant at 4.17×10^{22} water molecules per second or 75.4 ml per minute. So, in the (typically sub-ppm) range where the conductivity and concentration are linearly related, *the flow rate to achieve a 1% extraction remains unchanged, and independent of concentration.*

[0135] Finally, suppose that in either of the above two examples, one wished to measure lead concentrations in addition to arsenic concentrations. Because the flow rate is independent of concentration in the sub-ppm range, all that changes is that the arsenic atomic weight of 75 is replaced by the lead atomic weight of 207, which is an

207/75-fold increase. The lead ions will also be extracted at 10^{11} per second, and so a 1% extraction implies a 10^{13} per second supply. However, this means that one would need to flow $(207/75) \times 4.17 \times 10^{22}$ water molecules per second or 208 ml per minute. If the 4.17×10^{22} water molecules per second or 75.4 ml. per minute flow rate is maintained for purposes of a 1% arsenic extraction, then only $1/2.76 = .4\%$ of the lead ions will be extracted, which is certainly an acceptable number. Obviously, for multiple species, one would aim to have a maximum extraction of about 1% (though as noted above one can less preferably go as high as 5%) and would thus pick a flow rate giving an approximately 1% extraction rate to the lightest element to be detected, and thus giving somewhat lower extraction rates to the heavier elements to be detected.

[0136] The above values are provided as illustration of the flow rate required to limit the extraction to 1% of the impurity based on the conductivity of the matrix with 1 ppb concentration. At higher concentration levels, the conductivity and hence the extraction rate will be higher. However, there are proportionately more ions in the flow stream such that the percent of ions extracted will be the same at the same flow rate. Hence, the accuracy of the measurement is preserved over the entire range of impurity concentrations (typically sub-ppm) where the conductivity is essentially a linear function of the concentration. Therefore, the dynamic range of this instrument and measurement technique extends over many orders of magnitude of concentration levels from the sub-ppb to the ppm range to provide the same accuracy of measurement, and the variations introduced by differing atomic weights of different impurities species are easily dealt with.

[0137] If the extraction procedure is continued indefinitely, the impurities will continue to build up in the double layer. The amount of charge that the capacitor can hold is limited by its capacitance and the capacitor will eventually saturate. Once saturated, no net current will continue to flow. However, ion exchange will continue to take place with the high valence state ions replacing the low valence state ions in the double layer. This ion exchange process will distort the relative concentration of impurity ions trapped in the double layer, thereby distorting the measurement of concentrations in the flow stream.

[0138] To preserve the accuracy in interpretation of the data, it is important to avoid saturating the electrode surface. This is one important reason for requiring the electrode to be made of high surface density material such as NCC. Typically, such material is capable of holding 10 mg of foreign material per gm of native material, i.e. approximately 1% by weight, or 10,000 ppm. The detection sensitivity of the XRF technique is on the order of 1 ppm. Therefore, as long as the particular impurity concentration is greater than 0.01% of all the other impurities present, it would be in the detection range of the system. Stated simply, a system as disclosed herein employing NCC electrodes will not saturate until there is 10,000 times as much foreign matter in the NCC as is needed to achieve an XRF detection at a 1 ppm sensitivity.

[0139] This saturation can be observed by monitoring the leakage current. When the voltage is first applied to the electrodes, the ions in the immediate vicinity of the electrodes will be collected and appear as a measurable current. When these ions in the immediate vicinity are depleted, then the leakage current will gradually drop to an asymptotic value to reflect the rate at which ions in the flow stream can migrate under the effect of the electric field to the electrodes. This asymptotic leakage current, therefore, is a measure of the total concentration of dissolved ions in the matrix, as long as the electrodes are not saturated. As the electrode approaches saturation, the leakage current will gradually decrease and eventually vanish. Therefore, by monitoring the leakage current, the state of saturation of the electrodes can be monitored.

[0140] As stated above, once the electrode surfaces are fully saturated, no net electric current will flow. However, the impurities in the input flow stream 840 will continue to diffuse and can reach the electrodes where the higher valence state ions will replace the low valence state ions. By maintaining the flow after the electrode has saturated, the high valence state impurity ions will eventually build up to a sufficiently high level to become detectable. Since the concentration of the higher concentration impurities can be measured before saturation, the electrochemistry of the sample fluid can be understood, thereby allowing the ion exchange rate between the different impurity ions to be determined. This fact can be exploited using ultra-low trace element measuring means to detect and measure the concentration of ultra-low trace, high

valence ions in high conductivity fluids such as electrolytes, based on detecting the saturation state via the leakage current.

[0141] As discussed above, because extraction rates are kept low, the concentrations of impurities in the feed stream are not altered by more than a few percent at most. The ions are only captured until the analysis period is completed and then they are returned to the media from whence they came, as discussed further in connection with Figure 18. The invention disclosed herein requires only a small amount of NCC material in a single cell, because it is only used for detection and not removal. It can be configured to be controlled remotely over a telecommunications link, either directly, or by wireless means or via networks such as the Internet. The preferred system-specific operational user interface software is currently constructed in the LabViewTM architecture, a commercial program that allows for integrated control of hardware and software. LabView provides the means to set operating parameters and run the system via a virtual interface on a computer system connected to the system directly or remotely such as through the Internet or by wireless means. Alternatively, other terminal emulator programs can be employed, such as PC-anywhere and Timbuktu, or custom developed software to remotely control the system operation. Regardless of the means, because the operation of the present invention is capable of being fully automated, it can be controlled from remote locations. The remote operator would have all of the capabilities of an on-site operator such as system startup, emergency stop, test specifications, data analysis, data download and data review. The system can also be configured to operate automatically, reporting results of the analysis to remote locations. To prevent from unauthorized tempering, access to the operation of the system can be protected by password control. Obviously, the system would benefit from advances in remote computer communication technology and therefore the technology described above is a means of example and the present invention is not limited to the use thereof.

[0142] During or after the pre-concentration cycle described in Fig. 15, x-ray fluorescence is used to analyze the ions collected in the cell 112. Fig. 17 is a flowchart describing the measurement and analysis process in this system when this x-ray analysis is preformed *during the preconcentration cycle*. At block 960 the x-ray source emits the x-ray energy and it is directed towards the pre-concentration cell

As a monitoring instrument, the user can also preset certain predetermined threshold concentration values for the system to sound an alarm or trigger certain corrective or preventive actions automatically when the threshold value is exceeded (triggering means for triggering an action when the system detects that a concentration of at least one element of interest in said fluid has passed i.e., dropped below or exceeded a predetermined threshold concentration). For example, not limitation, an alarm can be employed for process fluids where a dissolved element is depleted by the process or builds up in the stream and needs to be automatically monitored and adjusted if the concentration falls outside of a specified range of values, and the action of either injection of concentrated material into the stream or dilution of the stream based on the alarm is utilized to maintain the specified values.

That is, the triggering means may trigger injecting at least one element (the detected element or another element) into the fluid, and / or it may trigger diluting at least one diluting fluid (more of the original fluid, or a different fluid) into the fluid.

[0144] As shown in Figure 18, at the end of the measurement cycle, the high surface area electrodes 218 and 220 are electrically shorted. As a result of employing such ionic release means for cleaning said preconcentration cell after use, the double layer disappears and the captured ions are released back into the output fluid flow 842 through the central flow interelectrode gap 222, such that they are returned to the media from whence they came. This resets the pre-concentration cell 100 to its initial state and thus completes the full cycle. The cell can thereafter be reused.

[0145] While the invention disclosed herein is intended to provide impurity measurements on site and can be installed on-line for continuous monitoring of the impurity concentrations in a flow stream, the pre-concentration cell can also function in a stand-alone mode to take samples at remote sites. This is *independent* of the X-ray and other testing, monitoring, detection, communications, and analysis equipment that further facilitates use of the pre-concentration cell. These remotely-gathered samples are then characterized by the X-ray source and detector/multi-channel analyzer system and procedures disclosed herein after the sampling is complete. Simply stated, samples can be collected in the cells themselves without all of the other equipment needed for testing and analysis, and the cells can then be moved to a different location for follow up testing and analysis.

[0146] As discussed in detail earlier, prior to its use for testing, each pre-concentration cell is first characterized (calibrated) to establish the background, sensitivity, and ion extraction rate. The pre-concentration cell, along with the voltage application, flow control, and leakage current monitoring hardware can be designed to be removable from the system to operate in a stand-alone mode. The pre-concentration cell can then be connected to any flow stream through standard flow connectors to sample the flow stream for a predetermined period of time based on estimated concentration of impurities in the flow stream determined from the leakage current observed. Upon completion of the sampling period, the outlet and intake valves on the pre-concentration cell are shut off, in that order, to trap some of the flow stream in the

cell before disconnecting from the source. The electrodes preferably remain charged by transportable voltage supply 385 to be shortly discussed, or the voltage may be turned off. In the absence of an applied voltage, the ions trapped in the double layer will, however, return to the fluid trapped in the cell, and a voltage will then need to be reapplied later on to re-form the double layer on the electrodes. The pre-concentration cell can then be inserted into the complete system x-ray system to measure the impurity concentrations in the cell.

[0147] A transportable preconcentration cell 101 suitable for this alternative operating method of sampling in an environment where x-ray source and/or detector equipment is not immediately available is shown in Fig. 19. In this method, the pre-concentration cell is configured for stand alone operation in order to collect the ionic particles at any desired location. After the preconcentration is complete, the pre-concentration cell is capped with the fluid inside and a voltage continuing to be applied, and transported to a laboratory or alternative environment equipped with the x-ray source, detector and analyzing system described above in Fig. 14.

[0148] Thus, as shown in Fig. 19, similarly to the preconcentration cell 100 of Fig. 13, *transportable* preconcentration cell 101 includes cell body 102, inlet flow port 104, inlet flow valve 380, outlet flow port 106, inlet flow slot 108, electrode assembly 112, fasteners 109, outlet flow slot 110, outlet flow valve 382, valve control 384, electric potential and current meter 386, and leakage current monitoring means 388. However, this transportable preconcentration cell 101 also comprises a transportable voltage supply 385 such as but not limited to one or more batteries or fuel cells, a control chip 387, a positive voltage cell lead 381, and a negative voltage cell lead 383. The transportable voltage supply 385 and control chip 387 are preferably embedded into cell body 102 and connected to leads 381 and 383 in order to minimize the need for excess space and to work integrally with the x-ray source and XRF detector in the same configuration as the pre-concentration cell 100 earlier shown in Figure 13. In a less preferred embodiment still within the scope of the invention, transportable voltage supply 385 and control chip 387 are maintained as separate modules attachable to the cell of Fig. 13 so as to supply the voltage.

[0149] The operation of transportable cell 101 follows the flow chart in Figure 20. First,

fluid flows through the cell and ions are captured in the preconcentration cell with voltage applied and maintained through voltage application means 389, which may comprise voltage from an external voltage source, or voltage from transportable voltage supply 385. Once preconcentration is complete, the cell is capped to keep the fluid and ions inside, and the voltage is either smoothly cut over from the external voltage source to the transportable voltage supply 385, or, if transportable voltage supply 385 was used all along, application of the voltage from transportable voltage supply 385 is simply maintained. This is all designated generally at block 962. This capping and continued application of a voltage maintains the ions in the double layer formation after preconcentration and throughout transport, and allows the cell to be transported to a laboratory with the necessary x-ray equipment shown in Fig. 14 to conduct immediate measurement and analysis without having to apply a voltage to re-entrain the ions in the electrodes via double layer formation.

[0150] It should be observed, while the cell of Fig. 19 is transportable due to the transportable voltage supply 385 which allows application a voltage across the electrodes during transport, that this cell can most certainly be used with the method described in Fig. 17 with the x-rays applied during preconcentration, wherein the voltage application means during preconcentration comprises transportable voltage supply 385 rather than some external voltage source. In other words, if desired, transportable voltage supply 385 can supply the voltage for preconcentration even when no transport is required, and can also be used for entrainment maintenance when transport is required.

[0151] Once preconcentration cell 101 has been transported to a laboratory with necessary x-ray equipment, it is exposed to x-rays via x-ray source 644, as designated by block 960. The x-ray fluorescence emissions from the preconcentration cell are then collected by detector 648, as designated by block 964. Thereafter, processing proceeds through blocks 964, 966, 968, 970, 972, and 974 in substantially the same manner as was earlier illustrated and described in connection with Fig. 17. When everything is complete, cleaning of the cell proceeds in the same manner as in Fig. 18.

[0152] At this point, we will discuss some further quantitative considerations pertaining

to this device, system and method, in order to provide a better idea of how all of the operational parameters relate to one another, to the samples being tested, and to the measurements that are desired.

[0153] As discussed above, the thickness of the electrode nearer to the source and detector should not exceed the optical depth of the photon to be detected, that is, this thickness should be less than or equal to .

[0154] The interelectrode gap, in turn, is determined by the pressure/flow rate conditions and the partial conductivity of the ions of interest. The Kapton windows can support about 1/10 atm. without excessive bowing (<100 microns). To limit the impurity extracted from the flow stream to around 1% of its initial value, with the applied voltage limited to approximately 1 Volt and a typical gap spacing of 1 mm, the flow rate needs to be at 1.5 mL/sec. or 90 mL/min. through a cell with a typical volume of 0.1 cm^3 . This is well within the flow rates achievable with 1/10 atm. inlet pressure for a cell volume defined by a 1 cm^2 window substantially matching the XRF interrogation spot area and a cell interelectrode gap of 0.1 cm.

[0155] Based on the known strength of Kapton and the thickness chosen, the window can support the pressure for flow rates significantly greater than the 90 mL/min required to maintain extraction well below 1%. Increasing the flow rate, however, does not increase the concentration time required to reach MDL. The concentration time required to reach MDL is given by:

$$t = \frac{N_i}{\Gamma_{ie}}$$

[0156] where N_i is the number of impurity ions captured by the electrodes required for detection, and Γ_{ie} is the rate these are extracted from the flow stream.

[0157] The number of impurity ions needed for detection is determined by the sensitivity of the instrument. Detection sensitivity, like impurity concentration, is usually defined in terms of mass ratio as:

$$S = \frac{N_i w_i}{V_c n_c w_c}$$

[0158] where V_c is the volume of the host material in the cell with number density n_c of

atomic weight w_c . Since most of the signal and background arise from the "upper" electrode closest to the "upper" window, V_c can be estimated as the volume of the first NCC electrode with ordinary surface area A and thickness d . If the electrode is filled with water, its number density and molecular density can be taken as that of water, $n_c = n_f = 3.3 \times 10^{22} \text{ \#}/\text{cm}^3$ and $w_c = w_f = 18$.

[0159] The extraction rate is previously given as:

$$\Gamma_{ie} = \frac{\sigma \Phi}{qd} A$$

[0160] Making the substitution into the time required to reach MDL, the resulting expression is:

$$t = \frac{w_f}{w_i} \frac{S \ell n_f q d}{\sigma \Phi} \propto \frac{S \ell}{\sigma} \propto \frac{S \ell}{C}$$

[0161] i.e., the concentration time required to reach MDL is directly proportional to the detection sensitivity of the instrument S and the thickness of the electrode d , and inversely proportional to the conductivity of the matrix σ . To understand the final proportionality, recall the important result that for low impurity concentration in the sub-ppm range, the conductivity is a linear function of the concentration that can be represented approximately as $\sigma \approx 2 \times 10^{-9} C / \Omega \text{-cm}$, where C is in units of ppb. Therefore, the concentration time required to reach MDL is inversely proportional to the concentration as is to be expected. Similarly, if the sensitivity of the system is poor, i.e. S is numerically larger, or if the electrode thickness is excessive to introduce more host material, the concentration time needed becomes longer. Note also that increasing the flow rate does not reduce the required concentration time, as has already been concluded, but does improve the accuracy of the measurement by reducing the change in concentration of the matrix, that is, by extracting a smaller percentage ϵ of contaminants. It is also worth noting that the time required to reach MDL is also independent of the ordinary surface area A covered by the electrode. This is because increasing the area increases the collection rate as well as the amount of host materials proportionately.

[0162]

As an example, consider again the concentration of 1 ppb arsenic in water in a cell with electrodes measuring 1 cm^2 by 100 micron thick separated by a distance of 2

mm with 1 Volt applied across them. The conductivity of the matrix with 1 ppb concentration is $2 \times 10^{-9} / \Omega \text{ -cm}$. Substituting these values into the expression for concentration time needed to reach MDL for an instrument with sensitivity $S = 10^{-6}$ (i.e. 1 ppm):

$$t = \frac{(18 \text{amu}) (10^{-6}) (2 \times 10^{-9} \text{cm}) (3.33 \times 10^{22} \text{cm}^{-3}) (1.6 \times 10^{-19} \text{Coulomb}) (10^{-1} \text{cm})}{(75 \text{amu}) (2 \times 10^{-9} / \Omega \text{ -cm}) (1 \text{ Volt})} = 1.28 \times 10^4 \text{ sec}$$

[0163] Therefore, to detect Arsenic in a matrix with a 1 ppb concentration may require 4 hours of concentration time. However, for 10 ppb, which is the current international standard for acceptance of As concentration in drinking water, the concentration time required is only one-tenth as long, i.e., approximately 25 min.

[0164] One question that arises particularly when one does not know *a priori* what concentration to expect is how long to run the cell? In many situations, the user will have a target MDL, for example, the user may wish to see if arsenic in a drinking water supply exceeds the EPA limit of 10ppb. Thus, the user runs the cell for the length of time necessary to detect 10ppb e.g., 25 minutes under the scenario described in the preceding paragraph and if there is nothing detected, then the user knows that the arsenic is below the EPA level. Thus, the concentration time here is based on the time required to accumulate a concentration up to a target MDL, that is, it is based on setting the impurity concentration C to a predetermined desired concentration detection level and then deducing the required concentration time accordingly. This timing control means can be automated, or can be manual.

[0165] Whether timing control is automated or manual, this means, in sum, that one can control how long the fluid flows through said ionic preconcentration cell based on setting an impurity concentration C , in a range where conductivity varies substantially linearly with concentration, to a predetermined desired concentration detection level and flowing the fluid for a time t given by:

$$t \propto \frac{S \ell}{\sigma} \propto \frac{S \ell}{C}$$

[0166] wherein S designates a sensitivity of the x-ray detection system; ℓ designates a thickness of the upper high surface area electrode; and σ designates a composite conductivity of the fluid.

[0167] In a different situation (with no target MDL), one could simply let the system run to saturation and then stop and obtain whatever sensitivity readings are yielded at saturation. In particular, there is a linear relationship between time and sensitivity which holds as long as saturation is not reached. The user would generally know which sensitivity is desired and would proceed on that basis. However, a monitoring of the leakage current alerts the user as to when saturation is approaching and the numbers might then not be as accurate as desired. Since the data is taken as a time series, this detection of the leakage current can then be used to "stop" the fluid flow (by the user or by the control system) to obtain the best/most reliable result, or more generally to control the period of time for which the fluid flows.

[0168] The function of the pre-concentration cell is to concentrate the impurities to facilitate their detection and quantitative measurement by identifying intensity peaks associated with the characteristic energy of the fluorescence photons in the spectrum. By maintaining a constant flow rate and keeping the extraction ratio small, the rate of concentration is constant and roughly independent of flow rate, as discussed above. Therefore, by reference to the established calibration and scaling values, the concentration in the sample can be related quantitatively to the concentration in the flow stream through lookup tables stored in the computer. For detection of different elements, the energy of the primary x-ray may be varied to maximize signal to noise ratio.

[0169] An example of the static operation of an XRF follows: In a region of the spectra where one would measure arsenic, the photon scattering into the continuum from the described window/water/NCC system (as an example, using a cell designed for atomic number 30, 4 cm² by about 0.2 cm thick) contributes about 0.5 counts/min/eV and the signal in the peak of the 1 ppm (0.8 μ grams) arsenic K-line is also about 0.5 counts/min/eV. After about 20 min the peak height is about ten counts as is the continuum. The fluctuation in the continuum is about three counts which yields a signal to noise of about 10/3 or three.

[0170] An example of the dynamic operation of an XRF follows: A 10ppb solution of As in water would contain 0.01microgram of As per cc of water and at a 120 cc/min (2 cc/sec) flow rate would be available at a rate of 1.2 micrograms/ minute. The

conductivity of such a solution is about 2×10^{-8} mho/cm. The resulting current is made up of both anions and cations so half of the current provides the ion deposition rate. With cell dimensions of $4\text{cm}^2 \times 0.15\text{ cm}$ (area * gap) and applying 1.5 volts, one would deposit 2.4×10^{12} ions/sec or 0.3 nanograms/sec or 18 nanograms/min. To deposit 0.8 micrograms would require 45 minutes and removes approximately 1.5% of the original As. The continuum background would grow linearly with time but both the signal and the amount of As deposited would grow linearly with time, hence the signal would grow quadratically with time. That is, $0.018\text{ micrograms/sec} \times t \times (0.5\text{ cnts/min/eV}/0.8\text{ micrograms}) \times t = 13.7\text{ counts/eV}$ for the signal at 35 minutes and the noise is $\text{sqrt}(0.5\text{ counts/min/eV} \times t) = 4.2$; thus $13.7/4.2 = \text{Signal/Noise} = 3.3$ at 35 minutes. This crude analysis uses peak to background for the signal to noise. A more sophisticated peak fitting/peak area analysis would show similar but better quantitative results.

[0171] Finally, once a cell has been used, it is necessary to clean the cell for its next use by purging the captured ions back into the fluid stream from whence they came. Various ionic release means for cleaning said preconcentration cell after use can be used to cleanse the cell automatically. They include removal of the applied voltage, temporary reversal of the applied voltage, or simply by shorting out the two electrodes. These techniques can be used in sequence or by themselves, depending on the structure of the electrode material and the level of impurity concentrations on their surfaces. The cleansing process returns all of the captured ions to the fluid stream. As the amount returned is no greater than the amount removed and this stream is usually of a smaller volume than the original stream, the transient increased concentration is negligible.

[0172] In an alternative embodiment of preconcentration cell 100 shown in Fig. 21, the inlet flow means comprises inlet flow slot 990 and the outlet flow means comprises outlet flow slot 992 respectively, but in a configuration differing somewhat from that of Fig. 1. In a further alternative embodiment shown in Fig. 22, the inlet flow means comprises an array of a plurality of inlet flow tubes 990 and the outlet flow means comprises an array of a plurality of outlet flow tubes 992. The array of inlet flow tubes 990 and the array of outlet flow tubes 992 both comprise a series of passageways of substantially circular cross section bored in the same plane as the

central flow interelectrode gap 222 which are all substantially parallel to each other with substantially equal spacing therebetween. The cross-sectional shape of inlet flow tubes 990 and outlet flow tubes 992 and their spacing and parallelism can be varied at will within the scope of this disclosure and its associated claims so long as these tubes permit the aqueous solution being tested to properly flow therethrough.

[0173] Aligned with the each of the inlet flow tubes 990 and the outlet flow tubes 992 of Fig. 22, but on diametrically opposite sides of inlet port 104 and outlet port 106, are a series of optional inlet flow channel access holes 994 and outlet flow channel access holes 996 respectively. These flow channel access holes 994 and 996 are an example of how to provide a debris cleaning means for cleaning debris from the inlet and outlet flow means, as necessary. Flow channel access holes 994 and 996 are closed and unused during the operation of the pre-concentration cell 100.

[0174] Additionally, inlet flow tubes 990 of the pre-concentration cell 100 optionally comprise flow turbulence enhancement means 850 for enhancing the turbulence of the input fluid flow 840, thereby inducing mixing of the flow stream to allow the impurities to be more uniformly extracted from the flow stream passing through the concentration cell. Such flow turbulence enhancement means can also accompany the embodiments utilizing flow slots as opposed to flow tubes. As shown in Fig. 23, the flow turbulence enhancement means 850 are arranged all along the array of input flow tubes 990 of the pre-concentration cell 100. Although the flow turbulence enhancement means 850 shown in Fig. 21 are illustrated as tabs placed on a diagonal at even intervals, they can actually be any shape and placed in any distribution that would be known to someone of ordinary skill for enhancing flow turbulence.

[0175] While only certain preferred features of the invention have been illustrated and described, many modifications, changes and substitutions will occur to those skilled in the art. It is, therefore, to be understood that this disclosure and its associated claims are intended to cover all such modifications and changes as fall within the true spirit of the invention.